

# PROFESSIONAL BREATH ALCOHOL ANALYSER

## JUPITER X




# Technical Data Sheet

Sensor	Fuel Cell
Operating System	Android 5.1
Wireless Communication	3G, WIFI
Working Temperature	-10°C to 50°C
Storage Temperature	-30°C to 70°C
Size	21.2cm X 8cm X 3.6cm
Weight	433g included battery
Battery	7.4V 2,000mAh NI-MH rechargeable battery
Battery Life	700 tests (250 tests with printout)
Response Time	5 seconds
Display LCD Size	4.3" TFT colour touch screen
Sampling System	Automatically or manually takes a lung sample
Optional Power	Supply 12 volt car adapter or AC 6 volt adapter
Camera	5 Megapixel 120° wide angle lens
Units of Measure	BrAC: mg/L, or other units BAC: grm%, mg/mL, mg/100ml, or other units
Detection Range	0.00% BAC to 0.40% BAC
Accuracy	0 - 0.10% BrAC ±5% 0.10 - 0.15% BrAC ±8% >0.15%: ±10%
Memory	100,000 tests records including all information
GPS	±5M
Connection Port	USB
Printer	Thermal printer



# Key Features

- Android 5.1
- Built-in Printer, GPS
- 3G/WIFI
- 5MP Camera
- provide evidential records by auto capturing while subject blowing
- 100,000 tests records stored in memory
- 4.3" Colour Touch Screen
- Mouthpiece can be automatically ejected



### Printout

```

Alcovisor-Jupiter X
Serial Number: 0000239
Record No: 0002
Date: 06.06.2018
Time: 17:38:15
Last Calibration Date:
Nov 19, 2018
Blow Test: 0.0 ug/100ml
Model Name:
Test Mode: Standard
Alcohol Content:
0.0 ug/100ml.

Subject's Name:
Marty
Vehicle Number:
A12345
License Number:
12345678
Police Officer's Name:
Paul
Badge Number:
A123
Department:
-----
GPS:
-----
Signature:
-----
Witness Signature:
-----

```

## Approvals

